




America

CERTIFICATE

No. U8 109610 0012 Rev. 00

Holder of Certificate: LPKF Laser & Electronics d.o.o.
 Polica 33
 4202 Naklo
 SLOVENIA

Certification Mark:



C US

Product: Laboratory Equipment

Model(s): Contac S4(v1.0 & v1.1)

Parameters:

Rated voltage:	110/230 VAC
Rated frequency:	50/60 Hz
Rated power:	750 kVA
Protection Class:	I

Tested according to:

- UL 61010-1:2012/R:2019-07
- CSA C22.2 No. 61010-1:2012/A1:2018-11
- UL 61010-2-010:2019
- CSA C22.2 No. 61010-2-010:2019

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713263180

Date, 2022-10-05

(Benedikt Pulver)